



NATIONAL INSTITUTE FOR STANDARDIZATION AND QUALITY
(Instituto Nacional de Normalização e Qualidade)
MOZAMBIQUE

CALIBRATION MEASUREMENT CAPABILITY - INNOQ (Non Accredited)		
LENGTH		
Tapes	$0 \leq L \leq 30 \text{ m}$	0.3 mm
calipers	$0 \leq L \leq 150 \text{ mm}$	0.011 mm
Micrometers	$0 \leq L \leq 100 \text{ mm}$	0.3 mm
DENSITY		
Higrometers	$0.1 \text{ g/cm}^3 \leq D \leq 1 \text{ g/cm}^3$	
TEMPERATURE INSTALATIONS		
Ovens, Mufles, Furnances, Climatic Chambers, Baths, Fridges and Refrigerators	-80 °C to -40 °C	1 °C
	-40 °C to 0 °C	0.45 °C
	0 °C to 125 °C	0.2 °C
	125 °C to 240 °C	0.36 °C
	240 °C to 420 °C	0.79 °C
	420 °C to 600 °C	1.2 °C
	600 °C to 1100 °C	2.2 °C