



Universität Stuttgart

Evaluation of the 5th SADCMET Water PT

Evaluation Workshop Kampala 2008

Dr.-Ing. Michael Koch
Institute for Sanitary Engineering, Water Quality and Solid Waste Management
Universität Stuttgart
Dep. Hydrochemie
Bandtäle 2
D-70569 Stuttgart
Tel.: +49 711 685 65444 / Fax: +49 711 685 67809
E-Mail: Michael.Koch@iswa.uni-stuttgart.de



Universität Stuttgart

Evaluation and Assessment

- according to same procedure as in the last rounds
 - assigned value from the formulation of the samples (with an uncertainty budget)
 - calculation of standard deviation using Algorithm A from ISO 13528
 - but! – limitation of the standard deviation (as ‘fitness for purpose’ requirement)

2 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Limits for standard deviation

parameter	std limit	parameter	std limit
sulphate	10 %	manganese	20 % / 12 %
chloride	10 %	aluminium	30 %
fluoride	12 %	lead	40 % / 25 %
nitrate	15 %	copper	20 %
phosphate	10 %	zinc	20 %
calcium	10 %	chromium	25 %
magnesium	10 %	nickel	25 %
sodium	10 %	cadmium	30 %
potassium	10 %	arsenic	30 %
iron	20 % / 12 %	cobalt	20 %

3 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Elimination of gross outliers

- Values < ref.-value/8 and > ref.-value*8 have been excluded before applying statistical procedures

4 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sulphate

Alg.A mean and ref.-value from weighings



Exp. uncertainty of the Alg.A mean is calculated according to ISO 13528: $U_{c_{\text{mean}}} = 2 \cdot u_{c_{\text{mean}}} = 2 \cdot 1,25 \cdot \frac{s_R}{\sqrt{n}}$


Exp. uncertainty of the ref.-value from an uncertainty budget



5 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sulphate

mean vs. ref.-value




Average recovery: 99.6%; in 2007: 103.6%; in 2006: 106.5%

6 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala




Sulphate calculated standard deviation and limit



No improvement over time

7 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sulphate Percentage non-satisfactory results





8 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Individual performance development

- For all labs participating in 2008 and in 2007 (or 2006)
- Calculation of the mean of the absolute values of z-scores of the 3 values
- Graphical display
 - all values
 - How many labs are
 - Consistently lower than 2.0 (good)
 - Consistently higher than 2.0 (bad)
 - Improving from > 2.0 to < 2.0
 - Getting worse from < 2.0 to > 2.0


9 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sulphate Individual performance development



10 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Sulphate 1



values: 35
removed: 1
mean: 19,66
ref.-value: 18,72
recovery: 105,0%
std: 5,016
rstd: 26,8%
std limit: 10%
upper limit: 22,47
lower limit: 14,98
too high: 9
too low: 5
outside limits: 14

11 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Sulphate 2



values: 36
removed: 1
mean: 43,61
ref.-value: 44,42
recovery: 98,2%
std: 11,227
rstd: 25,3%
std limit: 10%
upper limit: 53,30
lower limit: 35,54
too high: 6
too low: 8
outside limits: 14

12 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Sulphate 3



values:	36
removed:	1
mean:	70,00
ref.-value:	70,13
recovery:	99,8%
std:	16,184
rstd:	23,1%
std limit:	10%
upper limit:	84,15
lower limit:	56,10
too high:	6
too low:	7
outside limits:	13


13 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



14 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



15 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Sulphate

- Good agreement between means and ref.-values
- Standard deviation constantly higher than limits
- High portion of outliers for the gravimetical method

16 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chloride

Alg.A mean and ref.-value from weighings




Similar to 2006 data: means a little bit higher then ref.-values

17 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chloride


mean vs. ref.-value



Average recovery: 101.0%; in 2007: 102.4%; in 2006: 101.6%

18 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Chloride calculated standard deviation and limit



No improvement over time; standard deviation increases dramatically for low values

19 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Chloride Percentage non-satisfactory results



20 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chloride


Individual performance development



AQS Baden-Württemberg
iswa

21 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chloride 1




values:	41
removed:	1
mean:	13,86
ref.-value:	12,60
recovery:	110,0%
std:	3,857
rstd:	30,6%
std limit:	10%
upper limit:	15,12
lower limit:	10,08
too high:	12
too low:	4
outside limits:	16

AQS Baden-Württemberg
iswa


22 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chloride 2




values: 41
removed: 0
mean: 32,05
ref.-value: 31,55
recovery: 101,6%
std: 4,235
rstd: 13,4%
std limit: 10%
upper limit: 37,86
lower limit: 25,24
too high: 4
too low: 3
outside limits: 7

Chloride 3




values: 40
removed: 0
mean: 70,10
ref.-value: 69,67
recovery: 100,6%
std: 7,456
rstd: 10,7%
std limit: 10%
upper limit: 83,61
lower limit: 55,74
too high: 3
too low: 4
outside limits: 7

Used methods



25 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



26 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Chloride

- quite good agreement between mean and reference value
- standard deviation too high for low values
 - no improvement for higher values

27 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Fluoride


Alg.A mean and ref.-value from weighings



Means significantly higher than ref. Value, especially for low concentration

28 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Fluoride mean vs. ref.-value



Average recovery: 112.0%; in 2007: 98.2%; in 2006: 107.7%

29 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Fluoride calculated standard deviation and limit




For level 1 very high standard deviation – no improvement

30 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Fluoride



Percentage non-satisfactory results



31 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Fluoride

Individual performance development



32 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Fluoride 1




values: 20
removed: 4
mean: 0,40
ref.-value: 0,21
recovery: 195,1%
std: 0,263
rstd: 128,0%
std limit: 12%
upper limit: 0,25
lower limit: 0,16
too high: 11
too low: 2
outside limits: 13

No real consensus between the participants

33 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Fluoride 2



values: 21
removed: 0
mean: 1,09
ref.-value: 0,98
recovery: 110,8%
std: 0,229
rstd: 23,4%
std limit: 12%
upper limit: 1,22
lower limit: 0,75
too high: 5
too low: 0
outside limits: 5

34 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Fluoride 3



values: 21
 removed: 0
 mean: 1,46
 ref.-value: 1,32
 recovery: 110,7%
 std: 0,346
 rstd: 26,2%
 std limit: 12%
 upper limit: 1,64
 lower limit: 1,00
 too high: 6
 too low: 1
 outside limits: 7


35 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



36 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



37 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Fluoride

- mean values higher than reference values
- standard deviations too high, especially for low concentrations – no improvement
- colorimetric values not reliable (as in the last years!)

38 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Nitrate

Alg.A mean and ref.-value from weighings




Means slightly lower than reference values

39 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Nitrate


mean vs. ref.-value



Average recovery: 92,0%; in 2007: 85.9%; in 2006: 90.6%

40 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nitrate calculated standard deviation and limit



Highest standard deviation ever!!

41 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Nitrate Percentage non-satisfactory results



42 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nitrate

Individual performance development



43 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Nitrate 1




values:	37
removed:	4
mean:	9,58
ref.-value:	9,97
recovery:	96,1%
std:	4,559
rstd:	45,7%
std limit:	15%
upper limit:	12,96
lower limit:	6,98
too high:	7
too low:	9
outside limits:	16

most probably reported as $\text{NO}_3\text{-N}$ instead of NO_2^-

44 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nitrate 2



values: 36
removed: 4
mean: 26,77
ref.-value: 28,05
recovery: 95,4%
std: 10,001
rstd: 35,7%
std limit: 15%
upper limit: 36,47
lower limit: 19,64
too high: 4
too low: 11
outside limits: 15

45 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nitrate 3



values: 36
removed: 3
mean: 51,74
ref.-value: 56,82
recovery: 91,0%
std: 24,121
rstd: 42,4%
std limit: 15%
upper limit: 73,87
lower limit: 39,78
too high: 4
too low: 12
outside limits: 16

46 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Used methods



Many different methods hidden behind "colorimetric" – see separate presentation

47 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



48 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Nitrate

- some values obviously again reported in wrong units
- high number of outliers
- average quality is very bad! – no improvement!
- parameter needs more emphasis
- harmonization of methods?

49 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Phosphate


Alg.A mean and ref.-value from weighings



Means lower than reference values


50 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Phosphate mean vs. ref.-value



51 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Phosphate calculated standard deviation and limit




worse than in the last years

52 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Phosphate



Percentage non-satisfactory results



53 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Phosphate

Individual performance development



54 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Phosphate 1



values: 32
removed: 4
mean: 2,73
ref.-value: 3,20
recovery: 85,3%
std: 1,212
rstd: 37,8%
std limit: 10%
upper limit: 3,84
lower limit: 2,56
too high: 4
too low: 12
outside limits: 16

55 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Phosphate 2



values: 32
removed: 3
mean: 10,14
ref.-value: 11,34
recovery: 89,4%
std: 4,817
rstd: 42,5%
std limit: 10%
upper limit: 13,61
lower limit: 9,07
too high: 4
too low: 11
outside limits: 15


56 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Phosphate 3




57 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



58 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



59 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Phosphate

- values in wrong units
- very high standard deviation

60 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Calcium

Alg.A mean and ref.-value from weighings




consensus mean close to ref.value

61 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Calcium


mean vs. ref.-value



Average recovery: 101,6%; in 2007: 102.2%; in 2006: 97.2%

62 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Calcium calculated standard deviation and limit



worse than in the last years, especially for low concentrations

63 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Calcium Percentage non-satisfactory results



64 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Calcium

Individual performance development



65 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Calcium 1



values:	36
removed:	1
mean:	10,05
ref.-value:	8,35
recovery:	120,4%
std:	3,044
rstd:	36,5%
std limit:	10%
upper limit:	10,01
lower limit:	6,68
too high:	15
too low:	2
outside limits:	17


66 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Calcium 2




values: 36
removed: 0
mean: 25,80
ref.-value: 25,43
recovery: 101,5%
std: 5,375
rstd: 21,1%
std limit: 10%
upper limit: 30,51
lower limit: 20,34
too high: 7
too low: 6
outside limits: 13

Calcium 3




values: 35
removed: 0
mean: 61,02
ref.-value: 60,23
recovery: 101,3%
std: 7,832
rstd: 13,0%
std limit: 10%
upper limit: 72,28
lower limit: 48,19
too high: 4
too low: 3
outside limits: 7

Used methods



69 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



70 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Calcium

- mean values close to reference values
- standard deviations very high

71 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Magnesium


Alg.A mean and ref.-value from weighings



consensus mean close to ref.value

72 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Magnesium mean vs. ref.-value



Average recovery: 100,2%; in 2007: 101.7%; in 2006: 99.6%

73 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Magnesium calculated standard deviation and limit




Much worse than in the last years

74 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Magnesium



Percentage non-satisfactory results



75 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Magnesium

Individual performance development




76 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Magnesium 1




values: 33
removed: 2
mean: 12,35
ref.-value: 11,44
recovery: 108,0%
std: 4,303
rstd: 37,6%
std limit: 10%
upper limit: 13,73
lower limit: 9,15
too high: 9
too low: 5
outside limits: 14

Magnesium 2




values: 34
removed: 1
mean: 19,44
ref.-value: 18,85
recovery: 103,1%
std: 7,693
rstd: 40,8%
std limit: 10%
upper limit: 22,62
lower limit: 15,08
too high: 8
too low: 7
outside limits: 15

Magnesium 3




79 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



80 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



81 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Magnesium

- mean values around reference values
- standard deviations much too high
- titrimetric values not reliable

82 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sodium

Alg.A mean and ref.-value from weighings




similar to 2007

83 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sodium


mean vs. ref.-value



Average recovery: 100,3%, but too high for low conc. and low for high conc.
In 2007: 103.3%; in 2006: 104.4%

84 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Sodium calculated standard deviation and limit



Very high for lowest conc.

85 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Sodium Percentage non-satisfactory results



86 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Sodium

Individual performance development



87 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Sodium 1



values:	30
removed:	0
mean:	13,29
ref.-value:	10,10
recovery:	131,7%
std:	4,987
rstd:	49,4%
std limit:	10%
upper limit:	12,12
lower limit:	8,08
too high:	12
too low:	3
outside limits:	15


88 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Sodium 2




values: 30
removed: 1
mean: 46,54
ref.-value: 43,72
recovery: 106,4%
std: 8,489
rstd: 19,4%
std limit: 10%
upper limit: 52,46
lower limit: 34,98
too high: 6
too low: 3
outside limits: 9

Sodium 3




values: 30
removed: 1
mean: 79,06
ref.-value: 80,60
recovery: 98,1%
std: 17,567
rstd: 21,8%
std limit: 10%
upper limit: 96,72
lower limit: 64,48
too high: 4
too low: 8
outside limits: 12

Used methods



91 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



92 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Sodium

- consensus means close to ref.values
- standard deviations too high
- most probably calibration problems in some labs
- many too high values with FEP
- unreliable data with AAS

93 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Potassium


Alg.A mean and ref.-value from weighings



consensus mean close to ref.value


94 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Potassium mean vs. ref.-value



95 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Potassium calculated standard deviation and limit




Standard deviations worse than in 2007

96 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Potassium



Percentage non-satisfactory results



97 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Potassium

Individual performance development




98 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Potassium 1



Potassium 2




Potassium 3



values: 34
removed: 0
mean: 20,24
ref.-value: 20,72
recovery: 97,7%
std: 3,191
rstd: 15,4%
std limit: 10%
upper limit: 24,86
lower limit: 16,57
too high: 3
too low: 5
outside limits: 8


101 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



102 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



103 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Potassium

- Mean values close to reference values
- standard deviations higher than last year
- AAS values not reliable

104 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Iron


Alg.A mean and ref.-value from weighings



105 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Iron

mean vs. ref.-value




Average recovery: 99.9%; in 2007: 92.9%; in 2006: 88.0%
But for lowest level mean a bit too high

106 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Iron

calculated standard deviation and limit




higher than in last years

107 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Iron



Percentage non-satisfactory results



108 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Iron

Individual performance development



109 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Iron 1



values:	36
removed:	1
mean:	0,67
ref.-value:	0,54
recovery:	124,4%
std:	0,236
rstd:	43,8%
std limit:	20%
upper limit:	0,75
lower limit:	0,32
too high:	10
too low:	2
outside limits:	12

110 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Iron 2



values: 36
removed: 0
mean: 2,02
ref.-value: 2,03
recovery: 99,5%
std: 0,416
rstd: 20,4%
std limit: 12%
upper limit: 2,52
lower limit: 1,55
too high: 5
too low: 5
outside limits: 10

111 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Iron 3



values: 36
removed: 2
mean: 2,72
ref.-value: 2,74
recovery: 99,2%
std: 0,410
rstd: 15,0%
std limit: 12%
upper limit: 3,40
lower limit: 2,08
too high: 2
too low: 5
outside limits: 7


112 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



113 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



114 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Iron

- means close to reference values
- standard deviations too higher
- many outliers with colorimetric method

115 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Manganese

Alg.A mean and ref.-value from weighings




116 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Manganese mean vs. ref.-value



117 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Manganese calculated standard deviation and limit




higher than in 2007

118 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Manganese



Percentage non-satisfactory results



119 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Manganese

Individual performance development



120 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Manganese 1



values: 31
removed: 2
mean: 0,92
ref.-value: 0,90
recovery: 102,4%
std: 0,176
rstd: 19,6%
std limit: 20%
upper limit: 1,25
lower limit: 0,55
too high: 2
too low: 4
outside limits: 6

121 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Manganese 2



values: 31
removed: 1
mean: 1,50
ref.-value: 1,50
recovery: 100,1%
std: 0,232
rstd: 15,4%
std limit: 12%
upper limit: 1,86
lower limit: 1,14
too high: 2
too low: 5
outside limits: 7

122 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Manganese 3



values: 31
 removed: 3
 mean: 4,92
 ref.-value: 5,11
 recovery: 96,2%
 std: 0,680
 rstd: 13,3%
 std limit: 12%
 upper limit: 6,34
 lower limit: 3,88
 too high: 2
 too low: 7
 outside limits: 9


123 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



124 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



125 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Manganese

- mean values close to reference values
- standard deviation higher than in 2007

126 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Aluminium

Alg.A mean and ref.-value from weighings




consensus means a bit lower than ref.-values

127 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Aluminium


mean vs. ref.-value



Average recovery: 93.9%; in 2007: 96.1%; in 2006: 85.7%

128 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Aluminium calculated standard deviation and limit



Standard deviation higher

129 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Aluminium Percentage non-satisfactory results



130 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Aluminium

Individual performance development



131 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Aluminium 1



values:	15
removed:	1
mean:	0,78
ref.-value:	0,82
recovery:	95,0%
std:	0,375
rstd:	45,6%
std limit:	30%
upper limit:	1,32
lower limit:	0,33
too high:	1
too low:	2
outside limits:	3

132 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Aluminium 2



values: 15
removed: 1
mean: 2,14
ref.-value: 2,23
recovery: 95,7%
std: 0,703
rstd: 31,5%
std limit: 30%
upper limit: 3,57
lower limit: 0,89
too high: 0
too low: 3
outside limits: 3

133 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Aluminium 3



values: 16
removed: 2
mean: 4,10
ref.-value: 4,39
recovery: 93,4%
std: 0,762
rstd: 17,3%
std limit: 30%
upper limit: 5,91
lower limit: 2,87
too high: 0
too low: 4
outside limits: 4


134 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



135 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



136 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Aluminium

- small number of values
- mean values a bit below reference values

137 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Lead

Alg.A mean and ref.-value from weighings



138 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Lead mean vs. ref.-value



Average recovery: 103.7%; in 2007: 95.4%; in 2006: 95.6%

139 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Lead calculated standard deviation and limit




Higher than in last years

140 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Lead



Percentage non-satisfactory results



141 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Lead

Individual performance development



142 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Lead 1



values: 29
removed: 0
mean: 0,63
ref.-value: 0,58
recovery: 108,0%
std: 0,206
rstd: 35,5%
std limit: 25%
upper limit: 0,87
lower limit: 0,29
too high: 5
too low: 2
outside limits: 7

143 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Lead 2



values: 30
removed: 0
mean: 1,13
ref.-value: 1,12
recovery: 100,9%
std: 0,251
rstd: 22,4%
std limit: 25%
upper limit: 1,62
lower limit: 0,62
too high: 3
too low: 2
outside limits: 5

144 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Lead 3



values: 30
 removed: 1
 mean: 3,43
 ref.-value: 3,31
 recovery: 103,9%
 std: 0,477
 rstd: 14,4%
 std limit: 25%
 upper limit: 4,26
 lower limit: 2,35
 too high: 2
 too low: 4
 outside limits: 6


145 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



146 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



147 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Lead

- mean values around reference values
- standard deviation higher than in last years

148 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Copper


Alg.A mean and ref.-value from weighings



149 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Copper


mean vs. ref.-value



Average recovery: 95,1%, mainly caused by the highest level
in 2007: 97,5%; in 2006: 98,5%

150 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Copper calculated standard deviation and limit



Much higher than in previous PTs; 1st time higher than limit

151 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Copper Percentage non-satisfactory results



152 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Copper

Individual performance development



153 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Copper 1



values:	34
removed:	2
mean:	0,70
ref.-value:	0,70
recovery:	99,1%
std:	0,186
rstd:	26,5%
std limit:	20%
upper limit:	0,98
lower limit:	0,42
too high:	3
too low:	5
outside limits:	8

154 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Copper 2



values: 34
removed: 1
mean: 1,08
ref.-value: 1,09
recovery: 99,2%
std: 0,247
rstd: 22,7%
std limit: 20%
upper limit: 1,53
lower limit: 0,65
too high: 3
too low: 4
outside limits: 7

155 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Copper 3



values: 34
removed: 1
mean: 3,35
ref.-value: 3,54
recovery: 94,6%
std: 0,462
rstd: 13,0%
std limit: 20%
upper limit: 4,47
lower limit: 2,62
too high: 0
too low: 5
outside limits: 5


156 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



157 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



158 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Copper

- mean values in quite good agreement with reference values
- standard deviation higher than in previous years

159 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Zinc

Alg.A mean and ref.-value from weighings




consensus means slightly lower

160 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Zinc

mean vs. ref.-value




Average recovery: 95.5; in 2007: 93.0%; in 2006: 96.8%

161 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Zinc

calculated standard deviation and limit




Similar to previous years

162 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Zinc



Percentage non-satisfactory results



163 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Zinc

Individual performance development




164 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Zinc 1




values: 28
removed: 1
mean: 0,90
ref.-value: 0,88
recovery: 102,1%
std: 0,106
rstd: 12,0%
std limit: 20%
upper limit: 1,09
lower limit: 0,67
too high: 4
too low: 3
outside limits: 7

Zinc 2



values: 28
removed: 0
mean: 2,40
ref.-value: 2,48
recovery: 96,5%
std: 0,396
rstd: 15,9%
std limit: 20%
upper limit: 3,27
lower limit: 1,69
too high: 2
too low: 2
outside limits: 4


Zinc 3



values: 28
removed: 0
mean: 4,21
ref.-value: 4,43
recovery: 94,9%
std: 0,651
rstd: 14,7%
std limit: 20%
upper limit: 5,74
lower limit: 3,13
too high: 1
too low: 3
outside limits: 4


167 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



168 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



169 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Zinc

- mean values slightly lower than reference values
- standard deviation below limit

170 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chromium

Alg.A mean and ref.-value from weighings




Means slightly lower

171 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chromium


mean vs. ref.-value



Average recovery: 94,2%; in 2007: 100.1%; in 2006: 97.4%

172 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Chromium calculated standard deviation and limit



Similar to last years

173 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Chromium Percentage non-satisfactory results



174 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Chromium

Individual performance development



175 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Chromium 1



values:	28
removed:	6
mean:	0,48
ref.-value:	0,48
recovery:	99,2%
std:	0,126
rstd:	26,3%
std limit:	25%
upper limit:	0,72
lower limit:	0,24
too high:	4
too low:	6
outside limits:	10


176 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Chromium 2




values: 27
removed: 3
mean: 0,74
ref.-value: 0,77
recovery: 96,3%
std: 0,138
rstd: 17,8%
std limit: 25%
upper limit: 1,05
lower limit: 0,50
too high: 1
too low: 5
outside limits: 6

Chromium 3




values: 28
removed: 4
mean: 1,65
ref.-value: 1,77
recovery: 93,4%
std: 0,336
rstd: 19,0%
std limit: 25%
upper limit: 2,44
lower limit: 1,10
too high: 3
too low: 5
outside limits: 8

Used methods



179 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



180 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Chromium

- mean values slightly lower than reference values
- standard deviation around limit

181 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nickel

Alg.A mean and ref.-value from weighings




182 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Nickel mean vs. ref.-value



183 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Nickel calculated standard deviation and limit




Standard deviation for low level better than last year

184 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Nickel



Percentage non-satisfactory results



185 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nickel

Individual performance development



186 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nickel 1



values: 25
removed: 0
mean: 0,55
ref.-value: 0,54
recovery: 101,5%
std: 0,098
rstd: 18,1%
std limit: 25%
upper limit: 0,74
lower limit: 0,35
too high: 2
too low: 0
outside limits: 2

187 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nickel 2



values: 25
removed: 0
mean: 0,76
ref.-value: 0,77
recovery: 99,7%
std: 0,115
rstd: 15,0%
std limit: 25%
upper limit: 1,00
lower limit: 0,54
too high: 1
too low: 0
outside limits: 1

188 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Nickel 3



values:	25
removed:	0
mean:	3,06
ref.-value:	3,10
recovery:	98,5%
std:	0,365
rstd:	11,8%
std limit:	25%
upper limit:	3,83
lower limit:	2,37
too high:	1
too low:	1
outside limits:	2


189 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



190 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



191 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Nickel

- mean values in quite good agreement with reference values
- standard deviation ok

192 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Arsenic


Alg.A mean and ref.-value from weighings



193 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Arsenic


mean vs. ref.-value



Average recovery: 92.4%; mainly caused by the highest level
In 2007: 96.6%; in 2006: 111.2%

194 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Arsenic calculated standard deviation and limit



Similar to 2007

195 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala



Arsenic Percentage non-satisfactory results



196 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Arsenic

Individual performance development



197 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Arsenic 1



values:	7
removed:	0
mean:	0,11
ref.-value:	0,13
recovery:	85,7%
std:	0,033
rstd:	24,8%
std limit:	30%
upper limit:	0,20
lower limit:	0,07
too high:	0
too low:	0
outside limits:	0

198 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Arsenic 2



values: 8
removed: 0
mean: 0,40
ref.-value: 0,38
recovery: 105,8%
std: 0,147
rstd: 38,7%
std limit: 30%
upper limit: 0,61
lower limit: 0,15
too high: 1
too low: 0
outside limits: 1

199 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Arsenic 3



values: 9
removed: 0
mean: 0,50
ref.-value: 0,58
recovery: 87,0%
std: 0,165
rstd: 28,5%
std limit: 30%
upper limit: 0,91
lower limit: 0,25
too high: 0
too low: 0
outside limits: 0


200 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



201 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



202 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Arsenic

- low number of values
- standard deviation around limit

203 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cadmium


Alg.A mean and ref.-value from weighings



consensus means close to reference values

204 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Cadmium mean vs. ref.-value



Average recovery: 99.1%; in 2007: 96.4%; in 2006: 96.6%


205 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cadmium calculated standard deviation and limit





206 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cadmium Percentage non-satisfactory results




207 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cadmium Individual performance development



208 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Cadmium 1



values: 28
removed: 0
mean: 0,12
ref.-value: 0,12
recovery: 102,2%
std: 0,045
rstd: 37,3%
std limit: 30%
upper limit: 0,19
lower limit: 0,05
too high: 4
too low: 2
outside limits: 6

209 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Cadmium 2



values: 28
removed: 0
mean: 0,49
ref.-value: 0,49
recovery: 99,4%
std: 0,065
rstd: 13,1%
std limit: 30%
upper limit: 0,62
lower limit: 0,36
too high: 3
too low: 2
outside limits: 5

210 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Cadmium 3



values: 28
removed: 0
mean: 0,73
ref.-value: 0,73
recovery: 98,9%
std: 0,059
rstd: 8,0%
std limit: 30%
upper limit: 0,85
lower limit: 0,62
too high: 2
too low: 3
outside limits: 5


211 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



212 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods



213 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Summary Cadmium

- mean values in good agreement with reference values
- Standard deviation high for low concentration

214 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cobalt

Alg.A mean and ref.-value from weighings




consensus means close to reference values

215 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cobalt


mean vs. ref.-value



Average recovery: 99.8%


216 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cobalt calculated standard deviation and limit




217 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cobalt Percentage non-satisfactory results



218 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Cobalt 1



values: 19
removed: 1
mean: 0,98
ref.-value: 0,91
recovery: 108,4%
std: 0,175
rstd: 19,3%
std limit: 20%
upper limit: 1,25
lower limit: 0,56
too high: 3
too low: 0
outside limits: 3

219 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala


Cobalt 2



values: 19
removed: 0
mean: 2,63
ref.-value: 2,59
recovery: 101,7%
std: 0,420
rstd: 16,2%
std limit: 20%
upper limit: 3,43
lower limit: 1,75
too high: 2
too low: 0
outside limits: 2


220 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Cobalt 3




221 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Used methods



222 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Comparison of methods




223 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Summary Cobalt

- Consensus means close to reference values
- Standard deviation below limit


224 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Number of values per parameter




225 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Overview on participation




226 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Overview on participants' success



227 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Values not fit for purpose



228 Koch, M.: PT evaluation – SADCMET PT Workshop 2008 Kampala

Conclusion

- The PT Provider did a very good job
- The evaluation and assessment procedure is fit for the purpose
- The SADCMET Water PT is a good possibility for the participants to compare with peers and with stated fitness-for-purpose criteria
- The results of many laboratories are still not satisfactory or getting worse
- More emphasis should be put on corrective actions after unsatisfactory participation